

Yes See NA
or
Note DNI

3701-9-07 Ear Piercing Gun Standards

- (A) Training records for ear piercing gun
- (B) Disposable gloves shall be used and available
- (C) Ear piercing gun cleaned/disinfected after each use
- (D) Gun stored in covered container or cabinet
- (E) Patron notification of disinfection frequency/methods

3701-9-08 Sterilize & Disinfection Procedures

- (A) All non disposable equipment shall be cleaned and sterilized in the following manner:
- (1) Soaked in an enzymatic pre-cleaner
 - (2) Rinsed and patted dry
 - (3) Disassembled or placed in open position
 - (4) Visually inspected for cleanliness and damage
 - (5) Cleaned in tepid water and appropriate detergent
 - (6) Fully submerged in disinfectant per manufacturer
 - (7) Rinsed and patted dry
 - (8) Placed in ultrasonic unit filled with appropriate solution per manufacturer
 - (9) Rinsed and air dried
 - (10) Individually packed in sterilization pouches. Each pouch labeled with date of processing
 - (11) Sterilized in a steam sterilizer
 - Ultrasonic units and steam sterilizers used, cleaned, and maintained according to manufacturer. Records of maintenance kept for 2 yrs.

Yes See NA
or
Note DNI

(B) Monitor the function of sterilizers with the following:

- (1) Sterilization pouches with process indicator that changes color
- (2) Sterilization integrator used in each load
- (3) Weekly biological indicator tests submitted to lab
- (C) Documentation that indicators, integrators and biological tests were performed. Records are Maintained for 2 years and includes the following:
 - (1) Date and time the load was run
 - (2) Name of person who ran the load
 - (3) Results of integrator
 - (4) Report from lab on biological indicator test
- (C) Documentation kept in each patrons file for needles and instruments used on that patron.
- (D) New and replacement sterilizers shall be designed to sterilize hollow instruments and equipped with mechanical drying cycle
- (E) If wetness/moisture remains in/on pouches or if sterilizer malfunctions then instruments shall be considered contaminated and re-packaged/re-sterilized
- (F) Sterilized instruments remain in pouches until use
- (G) Malfunctioning sterilizer not used until repaired or replaced
- (H) Sterilized instruments stored in pouches, handled with gloves, stored in clean, dry, closed area. Re-sterilized if integrity of pouch is compromised.
- (I) Instruments re-sterilized after 1 year

Inspection Remarks

Facility utilizes non-disposable equipment.

- <> ensure spore tests are run weekly.
- <> integrators utilized and run in each autoclave load, then attached to the corresponding load cycle of the autoclave log. Ensure integrators continue to be used for each cycle.

- Sharps containers in place.

- Client Records complete.

- Artists' trainings complete.

Tony Schoenlein

Print Name of Inspector Completing Form

Inspector's Signature

September 8, 2017

Date